S.I.N. S.I.N.360

Get to know Implantat, the educational streaming of S.I.N.

implantat.global

CNPJ: 04.298.106/0001-74
SAC: 0800 770 8290
sin@sinimplantsystem.com

1-866-507-9315
sin360.us | sinimplantsystem.com

Visit our social networks

@sindentallUSA @sindentallUSA @sindentalusa



Versalis S

ENGINEERED LIKE ART. CRAFTED FOR VERSATILITY.


- Indicated for all bone types, mainly for low density bones, post-extraction socket and immediate and/or late loading.
- It can be used for all other clinical situations, as long as the clinical steps suggested in the drilling system are followed.
- Easy bicortical engagement, dual taper body and aggressive apical design ensure easier bicortical anchorage, even in challenging bone anatomy.
- Internal angulation: 16°.
- Available in heights: 8.5mm, 10.0mm, 11.5mm, 13.0mm, 15.0mm, 18.0mm and Long 20.0mm, 22.0mm, 24.0mm.

VERSALIS S INDICATIONS FOR CLINICAL USE:

- **3.5 mm** – Central incisors and lateral incisors (upper and lower)
- **3.8 mm** – Central incisors, lateral incisors (upper), canines and premolars
- **4.0 mm** – Central incisors (upper), canines, premolars and molars
- **4.3 mm** – Canines, premolars and molars
- **5.0 mm** – Molars
- **6.0 mm** – Molars
- **7.0 mm** – Molars

VERSALIS S LONG INDICATIONS FOR CLINICAL USE:

- **3.8 mm -** Anterior region
- **4.0 mm -** Anterior and posterior region
- **4.3 mm -** Posterior region



TECHNICAL SPECIFICATIONS:

- **1.5 mm infra-bone installation**
- **Initial drill speed:** 1200 rpm
- **Speed of the conical drills:** 800 rpm
- **Countersink speed:** 600-800 rpm
- **Insertion speed:** 20 to 40 rpm
- **Maximum torque:** 80 N.cm
- **Immediate loading*:** recommended torque from 45 to 80 N.cm
- **Includes cover screw of 2.0mm**

* Relative contraindication in patients with systemic or local problems and at the discretion of the professional.
* Check the availability of the products in your region.

PROSTHETIC FRAMEWORK

Versalis S

S.I.N.

PROSTHETIC FRAMEWORK VERSALIS S

*Check the availability of the products in your region.

◯ *Abutment Screw ■ *Squared Screw ⬢ *Hex driver
⚙ *Rotational component ⚙ *Anti-Rotational component

DIRECT SEQUENCE OVER THE IMPLANT

CEMENT-RETAINED PROSTHESIS

Single restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

TITANIUM HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CIM 3503C	3.5	2.0
CIM 3504C	3.5	4.0
CIM 3506C	3.5	6.0
CIM 4504C	4.5	4.0
CIM 4506C	4.5	6.0

TRANSFER

OPEN TRAY TRANSFER

CODE	DIAM. (mm)
TMAM 35C	3.5 mm
TMAM 45C	4.5 mm

CLOSED TRAY TRANSFER

CODE	DIAM. (mm)
TMAM 35C	3.5 mm
TMAM 45C	4.5 mm

PROVISIONALIZATION

TEMPORARY TITANIUM CYLINDER 20 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CPTM 3501 - H	3.5	1.0
CPTM 3502 - H	3.5	2.0
CPTM 3503 - H	3.5	3.0
CPTM 3504 - H	3.5	4.0
CPTM 4501 - H	4.5	1.0
CPTM 4502 - H	4.5	2.0
CPTM 4503 - H	4.5	3.0
CPTM 4504 - H	4.5	4.0

MODEL FABRICATION

ANALOG
ANMP 3800

FINAL PROSTHESIS

17° ANGLED CEMENTED ABUTMENT 20 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
AIAM 3502C-H	3.5	2.0
AIAM 3503C-H	3.5	3.0
AIAM 3504C-H	3.5	4.0
AIAM 3505C-H	3.5	5.0

CODE	DIAM. (mm)	HEIGHT (mm)
AIAM 4502C-H	4.5	2.0
AIAM 4503C-H	4.5	3.0
AIAM 4504C-H	4.5	4.0
AIAM 4505C-H	4.5	5.0

STRAIGHT CEMENTED ABUTMENT 20 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
AIMP 3502C-H	3.5	2.0
AIMP 3503C-H	3.5	3.0
AIMP 3504C-H	3.5	4.0
AIMP 3505C-H	3.5	5.0

CODE	DIAM. (mm)	HEIGHT (mm)
AIMP 4502C-H	4.5	2.0
AIMP 4503C-H	4.5	3.0
AIMP 4504C-H	4.5	4.0
AIMP 4505C-H	4.5	5.0

SCREWS

RETAINING SCREW
PT16
1.6 mm screw

LABORATORY-RELATED

LABORATORY SCREW
PTMAML 16
PTL 16

NOT APPLICABLE

DIGITAL

HEALING CAPS

PEEK HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CPCM 0504	5.0	4.0
CPCM 0804	8.0	4.0
CPCM 0508	5.0	8.0
CPCM 0808	8.0	8.0

TRANSFER

SCANNING JIG MT 16°
JIBWCM

SCANNING JIG MT 16°
JIBWCMC

PROVISIONALIZATION

TITANIUM INTERFACE MT 16° 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICMT 0504	0.5	4.0
ICMT 0506	0.5	6.0
ICMT 2004	2.0	4.0
ICMT 2006	2.0	6.0
ICMT 3004	3.0	4.0
ICMT 3006	3.0	6.0

TITANIUM INTERFACE MT 16° SIRONA 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICM 0804	0.8	4.0
ICM 2004	2.0	4.0

MODEL FABRICATION

MT 16° DIGITAL ANALOG
ADCM

FINAL PROSTHESIS

TITANIUM INTERFACE MT 16° 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICMT 0504	0.5	4.0
ICMT 0506	0.5	6.0
ICMT 2004	2.0	4.0
ICMT 2006	2.0	6.0
ICMT 3004	3.0	4.0
ICMT 3006	3.0	6.0

TITANIUM INTERFACE MT 16° SIRONA 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICM 0804	0.8	4.0
ICM 2004	2.0	4.0

SCREWS

RETAINING SCREW
PT16
1.6 mm screw

LABORATORY-RELATED

LABORATORY SCREW
PTMAML 16
PTL 16

NOT APPLICABLE

DIRECT SEQUENCE OVER THE IMPLANT

SCREW-RETAINED PROSTHESIS

Single restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

TITANIUM HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CIM 3503C	3.5	2.0
CIM 3504C	3.5	4.0
CIM 3506C	3.5	6.0
CIM 4503C	4.5	3.0
CIM 4504C	4.5	4.0
CIM 4506C	4.5	6.0

TRANSFER

OPEN TRAY TRANSFER

CODE	DIAM. (mm)
TMAM 35C	3.5 mm
TMAM 45C	4.5 mm

CLOSED TRAY TRANSFER

CODE	DIAM. (mm)
TMAM 35C	3.5 mm
TMAM 45C	4.5 mm

PROVISIONALIZATION

TEMPORARY TITANIUM CYLINDER 20 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CPTM 3501 - H	3.5	1.0
CPTM 3502 - H	3.5	2.0
CPTM 3503 - H	3.5	3.0
CPTM 3504 - H	3.5	4.0
CPTM 4501 - H	4.5	1.0
CPTM 4502 - H	4.5	2.0
CPTM 4503 - H	4.5	3.0
CPTM 4504 - H	4.5	4.0

MODEL FABRICATION

ANALOG
ANMP 3800

FINAL PROSTHESIS

CO-CR ABUTMENT (NO INTERNAL THREAD) 20 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
ECLAM 3501-H	3.5	1.0
ECLAM 3502-H	3.5	2.0
ECLAM 3503-H	3.5	3.0
ECLAM 3504-H	3.5	4.0
ECLAM 4501-H	4.5	1.0
ECLAM 4502-H	4.5	2.0
ECLAM 4503-H	4.5	3.0
ECLAM 4504-H	4.5	4.0

SCREWS

RETAINING SCREW
PT16
1.6 mm screw

LABORATORY-RELATED

LABORATORY SCREW
PTMAML 16
PTL 16

NOT APPLICABLE

DIGITAL

HEALING CAPS

PEEK HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CPCM 0504	5.0	4.0
CPCM 0804	8.0	4.0
CPCM 0508	5.0	8.0
CPCM 0808	8.0	8.0

TRANSFER

SCANNING JIG MT 16°
JIBWCM

SCANNING JIG MT 16°
JIBWCMC

PROVISIONALIZATION

TITANIUM INTERFACE MT 16° 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICMT 0504	0.5	4.0
ICMT 0506	0.5	6.0
ICMT 2004	2.0	4.0
ICMT 2006	2.0	6.0
ICMT 3004	3.0	4.0
ICMT 3006	3.0	6.0

TITANIUM INTERFACE MT 16° SIRONA 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICM 0804	0.8	4.0
ICM 2004	2.0	4.0

MODEL FABRICATION

MT 16 ° DIGITAL ANALOG
ADCM

FINAL PROSTHESIS

TITANIUM INTERFACE MT 16° 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICMT 0504	0.5	4.0
ICMT 0506	0.5	6.0
ICMT 2004	2.0	4.0
ICMT 2006	2.0	6.0
ICMT 3004	3.0	4.0
ICMT 3006	3.0	6.0

TITANIUM INTERFACE MT 16° SIRONA 20 N.cm

CODE	TRANSMUCOSAL HEIGHT (mm)	LENGTH (mm)
ICM 0804	0.8	4.0
ICM 2004	2.0	4.0

SCREWS

RETAINING SCREW 50 N.cm
PTMA 16

LABORATORY-RELATED

LABORATORY SCREW
PTMAML 16
PTL 16

NOT APPLICABLE

UNIVERSAL ABUTMENT

CEMENT-RETAINED PROSTHESIS

Single restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

TITANIUM HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CIM 3503C	3.5	2.0
CIM 3504C	3.5	4.0
CIM 3506C	3.5	6.0
CIM 4503C	4.5	3.0
CIM 4504C	4.5	4.0
CIM 4506C	4.5	6.0

PROSTHETIC INTERMEDIARY

CEMENTED UNIVERSAL ABUTMENT 20 N.cm

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AIM 3502C	3.3	4.0	1.0
AIM 3540C	3.3	4.0	2.0
AIM 3560C	3.3	4.0	3.0
AIM 3540C	3.3	4.0	4.0
AIM 3540C	3.3	4.0	5.0
AIM 3560C	3.3	6.0	1.0
AIM 3560C	3.3	6.0	2.0
AIM 3560C	3.3	6.0	3.0
AIM 3560C	3.3	6.0	4.0
AIM 3560C	3.3	6.0	5.0

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AIM 4502C	4.5	4.0	1.0
AIM 4540C	4.5	4.0	2.0
AIM 4560C	4.5	4.0	3.0
AIM 4540C	4.5	4.0	4.0
AIM 4540C	4.5	4.0	5.0
AIM 4560C	4.5	6.0	1.0
AIM 4560C	4.5	6.0	2.0
AIM 4560C	4.5	6.0	3.0
AIM 4560C	4.5	6.0	4.0
AIM 4560C	4.5	6.0	5.0

17° ANGLED CEMENTED UNIVERSAL ABUTMENT 50 N.cm *Use hexagonal driver 0.9 mm.

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AAIM 3370HC	3.3	1.5	4.0
AAIM 3374QC	3.3	2.5	4.0
AAIM 3374XC	3.3	3.5	4.0
AAIM 3376HC	3.3	1.5	6.0
AAIM 3376QC	3.3	2.5	6.0
AAIM 3376XC	3.3	3.5	6.0

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AAIM 4574HC	4.5	1.5	4.0
AAIM 4574QC	4.5	2.5	4.0
AAIM 4574XC	4.5	3.5	4.0
AAIM 4576HC	4.5	1.5	6.0
AAIM 4576QC	4.5	2.5	6.0
AAIM 4576XC	4.5	3.5	6.0

TRANSFER

POLYACETAL TRANSFER

CODE	DIAM. (mm)	HEIGHT (mm)	
TSIT 3340	3.3	4.0	yellow
TSIT 3360	3.3	6.0	blue
TSIT 4540	4.5	4.0	yellow
TSIT 4560	4.5	6.0	blue

UNIVERSAL ABUTMENT SCANNING JIG

CODE	DIAM. (mm)	HEIGHT (mm)
JBSIT 3340	3.3	4.0
JBSIT 3360	3.3	6.0
JBSIT 4540	4.5	4.0
JBSIT 4560	4.5	6.0

PROVISIONALIZATION

TEMPORARY ACRYLIC CYLINDER

CODE	DIAM. (mm)	HEIGHT (mm)
CPSIT 3340	3.3	4.0
CPSIT 3360	3.3	6.0
CPSIT 4540	4.5	4.0
CPSIT 4560	4.5	6.0

NOT APPLICABLE

MODEL FABRICATION

ANALOG

CODE	DIAM. (mm)	HEIGHT (mm)
ASIT 3340	3.3	4.0
ASIT 3360	3.3	6.0
ASIT 4540	4.5	4.0
ASIT 4560	4.5	6.0

DIGITAL ANALOG UNIVERSAL ABUT.

CODE	DIAM. (mm)	HEIGHT (mm)
ADUA 3340	3.3	4.0
ADUA 3360	3.3	6.0
ADUA 4540	4.5	4.0
ADUA 4560	4.5	6.0

FINAL PROSTHESIS

CALCINABLE POLYACETAL CYLINDER

CODE	DIAM. (mm)	HEIGHT (mm)
CCSIT 3340	3.3	4.0
CCSIT 3360	3.3	6.0
CCSIT 4540	4.5	4.0
CCSIT 4560	4.5	6.0

NOT APPLICABLE

SCREWS

RETAINING SCREW 50 N.cm
PTMA 16

LABORATORY-RELATED

NOT APPLICABLE

DIGITAL

HEALING CAPS

PEEK HEALING CAP 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
CPCM 0504	5.0	4.0
CPCM 0804	8.0	4.0
CPCM 0508	5.0	8.0
CPCM 0808	8.0	8.0

PROSTHETIC INTERMEDIARY

CEMENTED UNIVERSAL ABUTMENT 20 N.cm

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AIM 3502C	3.3	4.0	1.0
AIM 3540C	3.3	4.0	2.0
AIM 3560C	3.3	4.0	3.0
AIM 3540C	3.3	4.0	4.0
AIM 3540C	3.3	4.0	5.0
AIM 3560C	3.3	6.0	1.0
AIM 3560C	3.3	6.0	2.0
AIM 3560C	3.3	6.0	3.0
AIM 3560C	3.3	6.0	4.0
AIM 3560C	3.3	6.0	5.0

CODE	DIAM. (mm)	CEMENTATION LENGTH (mm)	TRANSMUCOSAL LENGTH (mm)
AIM 4540C	4.5	4.0	1.0
AIM 4540C	4.5	4.0	2.0
AIM 4560C	4.5	4.0	3.0
AIM 4540C	4.5	4.0	4.0
AIM 4540C	4.5	4.0	5.0
AIM 4560C	4.5	6.0	1.0
AIM 4560C	4.5	6.0	2.0
AIM 4560C	4.5	6.0	3.0
AIM 4560C	4.5	6.0	4.0
AIM 4560C	4.5	6.0	5.0

TRANSFER

POLYACETAL TRANSFER

CODE	DIAM. (mm)	HEIGHT (mm)	
TSIT 3340	3.3	4.0	yellow
TSIT 3360	3.3	6.0	blue
TSIT 4540	4.5	4.0	yellow
TSIT 4560	4.5	6.0	blue

UNIVERSAL ABUTMENT SCANNING JIG

CODE	DIAM. (mm)	HEIGHT (mm)
JBSIT 3340	3.3	4.0
JBSIT 3360	3.3	6.0
JBSIT 4540	4.5	4.0
JBSIT 4560	4.5	6.0

PROVISIONALIZATION

NOT APPLICABLE

MODEL FABRICATION

DIGITAL ANALOG UNIVERSAL ABUT.

CODE	DIAM. (mm)	HEIGHT (mm)
ADUA 3340	3.3	4.0
ADUA 3360	3.3	6.0
ADUA 4540	4.5	4.0
ADUA 4560	4.5	6.0

FINAL PROSTHESIS

NOT APPLICABLE

SCREWS

RETAINING SCREW 50 N.cm
PTMA 16

LABORATORY-RELATED

LABORATORY SCREW
PTMAL 01

POLISHING PROTECTOR
PTMA 01

NOT APPLICABLE

MULTIFUNCTIONAL ABUTMENT

SCREW-RETAINED PROSTHESIS

Single and multiple screw retained restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

ABUTMENT PROTECTOR 50 N.cm
PMA 48

PROSTHETIC INTERMEDIARY

MULTIFUNCTIONAL ABUTMENT 50 N.cm
*Use the 1.6 mm hex key from the prosthetic kit.

CODE	DIAM. (mm)	HEIGHT (mm)
AMCM 4803	4.8	3.0
AMCM 4802	4.8	2.0
AMCM 4803	4.8	3.0
AMCM 4804	4.8	4.0

TRANSFER

OPEN TRAY TRANSFER
TMAM 00
TMAM 06

CLOSED TRAY TRANSFER

CODE	DIAM. (mm)
TMFAM 00	
TMFAM 06	

PROVISIONALIZATION

TEMPORARY TITANIUM CYLINDER 50 N.cm
PTAM 00
PTAM 06

MODEL FABRICATION

ANALOG
ANAM

FINAL PROSTHESIS

CALCINABLE AND CO-CR CYLINDE 50 N.cm

CODE	
CCAM 00	Plastic
ICCAM 06	Plastic
CCRAM 00	Cobalt chrome
ICRAM 06	Cobalt chrome

SCREWS

RETAINING SCREW 50 N.cm

CODE	LENGTH (mm)
PRA 01	2.0

LABORATORY-RELATED

LABORATORY SCREW

CODE	DIAM. (mm)
PTMAL 01	1.4

POLISHING PROTECTOR
PRM 01

NOT APPLICABLE

DIGITAL

HEALING CAPS

ABUTMENT PROTECTOR 50 N.cm
PMA 48

PROSTHETIC INTERMEDIARY

MULTIFUNCTIONAL ABUTMENT 50 N.cm
*Use the 1.6 mm hex key from the prosthetic kit.

CODE	DIAM. (mm)	HEIGHT (mm)
AMCM 4803	4.8	3.0
AMCM 4802	4.8	2.0
AMCM 4803	4.8	3.0
AMCM 4804	4.8	4.0

TRANSFER

MULTIFUNCTIONAL ABUT SCANNING JIG
JBAM 00
JBAM 06

MULTIFUNCTIONAL ABUT SCANNING JIG
JBAM 00C
JBAM 06C

PROVISIONALIZATION

CHROME INTERFACE MULTIFUNCTIONAL ABUT. 50 N.cm

CODE	HEIGHT (mm)
IAMT 0400	4.0
IAMT 0600	6.0
IAMC 0400	4.0
IAMC 0600	6.0

TITANIUM INTERFACE MULTIFUNCTIONAL ABUT. 50 N.cm

CODE	HEIGHT (mm)
IAMC 0400	4.0
IAMC 0600	6.0
IAMT 0400	4.0
IAMT 0606	6.0

MODEL FABRICATION

DIGITAL ANALOG MULTIFUNCTIONAL ABUT.
ADAM

FINAL PROSTHESIS

CHROME INTERFACE MULTIFUNCTIONAL ABUT. 50 N.cm

CODE	HEIGHT (mm)
IAMT 0400	4.0
IAMT 0600	6.0
IAMC 0400	4.0
IAMC 0606	6.0

TITANIUM INTERFACE MULTIFUNCTIONAL ABUT. 50 N.cm

CODE	HEIGHT (mm)
IAMC 0400	4.0
IAMC 0600	6.0
IAMC 0406	4.0
IAMC 0606	6.0

SCREWS

RETAINING SCREW 50 N.cm

CODE	LENGTH (mm)
PRA 01	2.0

LABORATORY-RELATED

LABORATORY SCREW

CODE	DIAM. (mm)
PL 1405	1.4 mm
PTMA 13-1	1.4 mm

Short
Long

POLISHING PROTECTOR
PRM 01

NOT APPLICABLE

MULTI-UNIT ABUTMENT

SCREW-RETAINED PROSTHESIS

Multiple screw retained restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

ABUTMENT PROTECTOR 50 N.cm
PMA 4855

PROSTHETIC INTERMEDIARY

STRAIGHT MULTI-UNIT ABUTMENT 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
MAM 4801-C	4.8	1.0
MAM 4802-C	4.8	2.0
MAM 4803-C	4.8	3.0
MAM 4804-C	4.8	4.0

TRANSFER

OPEN TRAY TRANSFER
TMAM 4800

CLOSED TRAY TRANSFER
TMFM 4800

PROVISIONALIZATION

TEMPORARY TITANIUM CYLINDER 50 N.cm
PTM 4800-3
PTM 4800-2

For straight multi-unit
For straight multi-unit
Suitable for laser welding
For angled multi-unit

FINAL PROSTHESIS

CALCINABLE AND CO-CR CYLINDE 50 N.cm

CODE	
CCM 00-3	Plastic/For straight multi-unit
CLEM 4800-3	Cobalt chrome/For straight multi-unit
ICPM 4800-3	Plastic/For angled multi-unit
CLEM 4800-2	Cobalt chrome/For angled multi-unit

SCREWS

RETAINING SCREW 50 N.cm

CODE	LENGTH (mm)
PRH 20	2.0

For angled multi-unit

LABORATORY-RELATED

LABORATORY SCREW

CODE	DIAM. (mm)
PL 1405	1.4 mm
PTMA 13-1	1.4 mm

Short
Long

POLISHING PROTECTOR
PRM 01

NOT APPLICABLE

DIGITAL

HEALING CAPS

ABUTMENT PROTECTOR 50 N.cm
PMA 4855

PROSTHETIC INTERMEDIARY

INDEXED ANGLED MULTI-UNIT ABUTMENT 20 N.cm
*Use hexagonal driver 1.2 mm.

CODE	DIAM. (mm)	HEIGHT (mm)	ANG.
MAAM 4803	4.8	3.0	17°
MAAM 4803R	4.8	3.0	17°
MAAM 4804R	4.8	4.0	17°
MAAM 4803	4.8	3.0	30°
MAAM 4803R	4.8	3.0	30°
MAAM 4804R	4.8	4.0	30°

TRANSFER

MULTI-UNIT ABUT SCANNING JIG
JBMA

MULTI-UNIT ABUT SCANNING JIG
JBMAC

PROVISIONALIZATION

CHROME INTERFACE MULTI-UNIT ABUT. 50 N.cm

CODE	HEIGHT (mm)
IMAC 04	4.0
IMAC 06	6.0

TITANIUM INTERFACE MULTI-UNIT ABUT 50 N.cm

CODE	HEIGHT (mm)
IMAT 04	4.0
IMAT 06	6.0

MODEL FABRICATION

DIGITAL ANALOG MULTI-UNIT ABUT.
ADMA

FINAL PROSTHESIS

CHROME INTERFACE MULTI-UNIT ABUT. 50 N.cm

CODE	HEIGHT (mm)
IMAC 04	4.0
IMAC 06	6.0

TITANIUM INTERFACE MULTI-UNIT ABUT 50 N.cm

CODE	HEIGHT (mm)
IMAT 04	4.0
IMAT 06	6.0

SCREWS

RETAINING SCREW PACK 4 50 N.cm

CODE	LENGTH (mm)
PRH 30	3.0

LABORATORY-RELATED

NOT APPLICABLE

MICRO MULTI-UNIT ABUTMENT

SCREW-RETAINED PROSTHESIS

Single and Multiple screw retained restorations

TYPE OF PROSTHESIS

ANALOG

HEALING CAPS

ABUTMENT PROTECTOR 50 N.cm
PMA 33

PROSTHETIC INTERMEDIARY

MICRO MULTI-UNIT ABUTMENT 50 N.cm

CODE	DIAM. (mm)	HEIGHT (mm)
MAM 3301	3.5	1.0
MAM 3302	3.5	2.0
MAM 3303	3.5	3.0
MAM 3304	3.5	4.0

TRANSFER

OPEN TRAY TRANSFER
TMM 33
TMM 3306

CLOSED TRAY TRANSFER
TMMF 33
TMMF 3306

MICRO MULTI-UNIT ABUT SCANNING JIG
JBMMA
JBMMA 06

MICRO MULTI-UNIT ABUT SCANNING JIG
JBMMAC
JBMMA 06C

PROVISIONALIZATION

TEMPORARY TITANIUM CYLINDER 50 N.cm
CPMT 33
CPMT 3306

CHORME INTERFACE MICRO MULTI-UNIT ABUT 50 N.cm

CODE	HEIGHT (mm)
IMMAC 04	4.0
IMMAC 06	6.0
IMMAC 0406	4.0
IMMAC 0606	6.0

TITANIUM INTERFACE MICRO MULTI-UNIT ABUT. 50 N.cm

CODE	HEIGHT (mm)
IMMAT 04	4.0
IMMAT 06	6.0
IMMAT 0406	4.0
IMMAT 0606	6.0

MODEL FABRICATION

ANALOG
AMMA 33

DIGITAL ANALOG MICRO MULTI-UNIT ABUT.
ADMMA

FINAL PROSTHESIS

CALCINABLE AND CO-CR CYLINDE 50 N.cm

CODE	
CPMC 33	Plastic
CPMM 33	Plastic
ICPMC 3306	Cobalt chrome
ICPMM 3306	Cobalt chrome

CHORME INTERFACE MICRO MULTI-UNIT ABUT 50 N.cm

CODE	HEIGHT (mm)
IMMAC 04	4.0
IMMAC 06	6.0
IMMAC 0406	4.0
IMMAC 0606	6.0

TITANIUM INTERFACE MICRO MULTI-UNIT ABUT 50 N.cm

CODE	HEIGHT (mm)
IMMAT 04	4.0
IMMAT 06	6.0
IMMAT 0406	4.0
IMMAT 0606	6.0

SCREWS

RETAINING SCREW 50 N.cm
PRH 3035

LABORATORY-RELATED

LABORATORY SCREW
PTMAML 56

POLISHING PROTECTOR
PRMM 33
PRMM 3306

NOT APPLICABLE